

E.U. WEIGHBRIDGE VERIFICATION CERTIFICATE – No: 5443

CUSTOMER: Ash Waste	TYPE APPROVAL CERTIFICATE No: LNE-14588 Rev.3
SITE ADDRESS: Ashfield House, Dunkirk Way Dunkirk, Cheshire CH1 6LZ	INDICATOR TYPE: X222-B (i300)
MANUFACTURER: Precia Molen	INDICATOR TEST CERTIFICATE No: LNE-25599-Rev 1
WEIGHBRIDGE TYPE: VS400	INDICATOR SERIAL No: 10F763627
PLATFORM SIZE: 9m x 3m	MODULE TEST CERTIFICATE No: LNE TC 05-02 Rev 1
EST. PLATFORM WEIGHT: 4,500 kg	MAXIMUM CAPACITY: 40,000 kg
WEIGHBRIDGE LOCATION: Main Yard Weighbridge	DIVISION (e): 20 kg
INITIAL VERIFICATION: <input type="checkbox"/>	LOADCELL TEST CERTIFICATE No: NMi TC5404 Rev3
	PRINTER SERIAL No: N/A
	TARE FACILITY: Disabled
RE-VERIFICATION: <input checked="" type="checkbox"/>	

Repeatability Test - MPE 0.3e

	Indication (HR)	Indication (HR)	Indication (HR)
Loaded WTU	39380	39382	39382
	PASS	✓	FAILED

Accuracy of Zero, Linearity/Hysteresis, & Comparison Tests

Approximate Test Interval	MPE	Actual Load	Indication Up (HR)	Error	Corrective Error	Indication Down (HR)	Error	Corrective Error	Comparison
Zero	0.25	0	2	0.1e	0.1e	2	0.1e	0e	
2e	0.5	40	42	0.1e	0e	44	0.2e	0.1e	
20e	0.5	440	442	0.1e	0e	444	0.2e	0.1e	
500e	1.0	10040	10044	0.2e	0.1e	10044	0.2e	0.1e	
1000e	1.0	20040	20044	0.2e	0.1e				
Substitute Load	1.0	19376	19380	0.2e	0.1e	19380	0.2e	0.1e	No
1250e	1.0	24376	24380	0.2e	0.1e	24382	0.3e	0.2e	
1500e	1.0	29376	29380	0.2e	0.1e	29380	0.2e	0.1e	
2000e	1.0	39376	39382	0.3e	0.2e				
Substitute Load									
2200e									
PASSED ✓					FAILED				

Loadcell Positions

2	3
1	4

DWI

FLT Substitute Load
3,822 kg

Comparison Test

Printer	
Remote Display	
PC	
Other	
Pass	Fail

Eccentric Load Test – Up to 500e MPE = 0.5e, over 500e MPE = 1.0e

Position	1	2	3	4	5	6	7	8	9	10
Test Load	14040	14040	14040	14040						
Indication (HR)	14044	14044	14046	14044						
Error	0.2e	0.2e	0.3e	0.2e						
PASSED ✓					FAILED					

ENVIRONMENTAL CONDITIONS: Dry	AUTHORISED ENGINEER: C.Swinburne
TEST WEIGHTS USED: Precia-Molen UK Ltd	CERTIFICATE NO: W18034 & 70591903
VERIFICATION DATE: 14-08-18	NEXT AFC DATE : 14-08-19

Signature:  Date:03/09/2018.....

The EU Weighbridge Verification Certificate complies with the requirements of NAWI directive 2014/31/EU and EN45501:2015. The tests are also in compliance with the UK Weighing Federation Code of Practice for Non Automatic Weighing Instruments.
EU Weighbridge Verification Certificate. – Issue: 16.1 Dated: 03/07/2018