

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	ASH NORTH WEST
SITE REFERENCE:	4518415
SITE ADDRESS:	DUNKIRK
POST CODE:	CH16LZ
WEIGHBRIDGE LOCATION:	ON SITE
WEIGHBRIDGE SIZE:	9 X 3

REPORT NO.:	1
INDICATOR TYPE:	1 300
INDICATOR SERIAL NO.:	10F763627
MAXIMUM CAPACITY (kg):	40,000
DIVISION 'e' (kg):	20

Weather Conditions: SUNNY

REPEATABILITY TEST

Direction of Travel →
←

Indicator Reading: 24,160 kg
Weighbridge Viewed from Indicator Position
Indicator Reading: 24,160 kg

Error Allowed
+/-2e or +/-40kg
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	15,000
-------------------	--------

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
4,160	15,000	15,000	15,000	15,000	0
	ERROR (e)	error +/-2dvs or 40kg		error +/-2dvs or 40kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Added weight (kg)	Allowed Error
0	0	0	0.0	Pass		error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass		error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass		error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass		error +/-2dvs or 40kg
Substitute Load	19,160	19,160	0.0	Pass		error +/-2dvs or 40kg
28,600	24,160	24,160	0.0	Pass	5,000	error +/-2dvs or 40kg
38,600	34,160	34,160	0.0	Pass	10,000	error +/-2dvs or 40kg
43,600	39,160	39,160	0.0	Pass	5,000	error +/-2dvs or 40kg
44,000	39,560	39,560	0.0	Pass	400	error +/-2dvs or 40kg

MAINT LOG UPDATED: No

MAXIMUM RANGE TESTED: No

TEST EQUIPMENT USED: Test Weights Traceable to National Standard

CALIBRATION DATE: 29TH JULY 2019

AUTHORISED ENGINEER:

Leslie Jarvis

CERTIFICATE NUMBER:

W18226

NEXT CALIBRATION DATE:

28 July 2020

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Print name	J C THORWICK	Date	29-7-2019
Signature		Engineer's Signature	

We Certify that the Weighbridge Calibration check complies to the current Weights & Measures criteria and the calibration procedure is carried out in accordance with the Weighing Federation's Code of practice.

(v13LH)