

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	ALANS SKIPS HIRE
SITE REFERENCE:	4640412
SITE ADDRESS:	BROUGHTON MILLS ROAD BRETTON CHESTER
POST CODE:	CH4 0BY
WEIGHBRIDGE LOCATION:	MAIN WEIGHBRIDGE
WEIGHBRIDGE SIZE:	17x3

SAP REPORT NO.:	4640412
INDICATOR TYPE:	X201
INDICATOR SERIAL NO.:	28F630403
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	20
Weather Conditions	RAIN

REPEATABILITY TEST

Direction of Travel	→	Indicator Reading: 43,700 kg	Error Allowed +/-2dvs or +/-40kg
		Weighbridge Viewed from Indicator Position	
	←	Indicator Reading: 43,720 kg	Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	10,260
-------------------	--------

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
4,260		10,260	10,260	10,260	0
	ERROR (e)	error +/-2dvs or 40kg		error +/-2dvs or 40kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Allowed Error
0	0	0	0.0	Pass	error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass	error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass	error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass	error +/-2dvs or 40kg
Substitute Load	19,460	19,460	0.0	Pass	error +/-2dvs or 40kg
25,000	24,460	24,440	-1.0	Pass	error +/-2dvs or 40kg
35,000	34,460	34,460	0.0	Pass	error +/-2dvs or 40kg
40,000	39,460	39,460	0.0	Pass	error +/-2dvs or 40kg
Substitute Load	38,700	38,700	0.0	Pass	error +/-2dvs or 40kg
44,000	43,700	43,700	0.0	Pass	error +/-3dvs or 60kg

MAINT LOG UPDATED: NO

MAXIMUM RANGE TESTED: NO

TEST EQUIPMENT USED: Test Weights Traceable to National Standards

CALIBRATION DATE: 06 October 2020

AUTHORISED ENGINEER:

CERTIFICATE NUMBER:

NEXT CALIBRATION DATE:

Andrzej Frejer

W19203; M19204

05 October 2021

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Print name		Date	06/10/2020
Signature		Engineer's Signature	

We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.

v14. May 2016