

CERTIFICATE OF CALIBRATION

Certificate No:	WS3966	Date:	28/01/21
------------------------	---------------	--------------	-----------------

Company Name:	ASH Waste
Address:	Redwither Road, Wrexham Ind Est, LL13 9RD.


Machine Make:	Weighbridge Services	Location:	Yard
Serial No:	04514815	Minimum Load:	400kg
Weight Display:	3590ET	Capacity:	40000 kg

Class M1 Test Weights Used:	Certificate Number TR11150	2 x 200kg Block Weight
	Certificate Number TR11149	20 x 1000kg Block weight

Next Test Date:	January 2022
------------------------	---------------------

Weight and Measures tolerances for use with a Class III Machine

0-499 Divisions = +/- 1 Division 500-1999 Divisions = +/- 2 Divisions 2000-3000 Divisions = +/- 3 Divisions

Off Centre Position Loading		
		
Load Position	Weight Applied	Weight Reading
1	10,000kg	10,000kg
2	10,000kg	10,000kg
3	10,000kg	10,000kg
4	10,000kg	10,000kg
5	10,000kg	10,000kg
6	10,000kg	10,000kg

Weighbridge Calibration		
Load Applied	Pre Adjustment	Post Adjustment
0kg	N/A	0kg
400kg	N/A	400kg
10,000kg	N/A	10,000kg
20,000kg	N/A	20,000kg
35,960kg	N/A	35,960kg
45,960kg	N/A	45,960kg